

COMPANY SURPLUS EQUIPMENT: CVD SYSTEM



SYSTEM SPECIFICATIONS:

- 316 Stainless steel chamber: (08 inch diameter; 10 inch height, chamber (316 SS) complete with water cooling
- Hewlett Packard (HP) 6269B DC Power supply (0-40V, 0-50A)
- Hastings vacuum gauge readout unit(model EVT-5A-J) and gauge head (reading pressure in microns of mercury)
- Edwards Pirani gauge – Model PRL 10
- Edwards pirani penning 1005
- MKS model 247B 4-channel readout unit
- Quartz viewing port
- Three electrode electrical feedthroughs
- Thermocouple (cap. To measure from 0- above 1000 OC)

CONTACT

Dr N. Ali

CNC Coatings,
Regal House,
Miall Street,
Rochdale,
OL11 1HY, UK



E-mail:
n.ali@cnc-coatings.co.uk

Tel:
[+44 \(0\) 1706711007](tel:+44(0)1706711007)